



Title: **NADMED Lot 0003 and 0004**

Document type: **Certificate of Analysis**

Document ID: **COA-1**

Author: Riikka Äänismaa

Owner: Riikka Äänismaa

Approver(s):
Jana Buzkova
approved at 2023-03-10 13:52 (UTC +0200)
Liliya Euro
approved at 2023-03-10 14:16 (UTC +0200)

Approval date: 2023-03-10

Effective date: 2023-03-10

Next periodic
review date: 2025-03-10

1. CERTIFICATE OF ANALYSIS

Product: Q-NADMED blood kit, Catalog number IVD_01

Lot: NADMED - 0003 and 0004

Specification: Performance evaluation upon manufacturing


Tested parameters:

1. UV-Vis spectra of NAD⁺ and/or NADH standard stocks
2. Performance of the Standards in the assay
3. Volume of individual components in the kit

Status: All parameters are within reference range


Date: 24.02.2023

Approved by Liliya Euro, PhD



2. Summary of UV-Vis spectroscopy analysis of kit standards

Summary of UV-Vis spectroscopy analysis of kit standards
Purpose: measurement of compound concentration in NAD ⁺ and NADH standard stocks




Standard: 1 mM NAD+
 Extinction coefficient at 260 nm: 18 mM*cm⁻¹
 Dilution: x100 with water to 10 µM
 Measurement: 1 cm quartz cuvette
 Instrument: Shimadzu UV-2401pc
 Absorbance at 260 nm: 0.180 Optical Units
 Theoretical value for absorbance of 10µM NAD+: 0.180
 Accepted range for absorbance measurement based on instrument characteristics: 0.178 - 0.185

Standard: 1 mM NADH
 Extinction coefficient at 340 nm: 6.22 mM*cm⁻¹
 Dilution: x100 with water to 10 µM
 Measurement: 1 cm quartz cuvette
 Instrument: Shimadzu UV-2401pc
 Absorbance at 260 nm: 0.061 Optical Units
 Theoretical value for absorbance of 10µM NADH: 0.065
 Accepted range for absorbance measurement based on instrument characteristics: 0.059-0.066

3. Summary of Standard performance in the assays

Assay	Absorbance at 573 nm	Standard curve fitting																		
NAD+ Standards	Reaction time - 4 min																			
	<table border="1"> <thead> <tr> <th>NAD+, uM</th> <th>1 meas</th> <th>2 meas</th> </tr> </thead> <tbody> <tr> <td>0</td> <td>0,054</td> <td>0,054</td> </tr> <tr> <td>1</td> <td>0,201</td> <td>0,195</td> </tr> <tr> <td>2</td> <td>0,329</td> <td>0,34</td> </tr> <tr> <td>3</td> <td>0,487</td> <td>0,467</td> </tr> <tr> <td>5</td> <td>0,787</td> <td>0,744</td> </tr> </tbody> </table>		NAD+, uM	1 meas	2 meas	0	0,054	0,054	1	0,201	0,195	2	0,329	0,34	3	0,487	0,467	5	0,787	0,744
	NAD+, uM		1 meas	2 meas																
	0		0,054	0,054																
	1		0,201	0,195																
	2		0,329	0,34																
3	0,487	0,467																		
5	0,787	0,744																		
NADH Standards	Reaction time - 6 min																			



4. Criteria for acceptance assay performance results

Parameter	Reference values	Quality Control
Absorbance range for 0 - 5 µM NAD+ Assay Standards (assay time 4 min)	0.04 - 0.800 Accepted variation between replicates - 0.05 Optical Units	passed
Absorbance range for 0 - 1 µM NADH Assay Standards (assay time 6 min)	0.04 - 0.260 Accepted variation between replicates - 0.05 Optical Units	passed
R ² of liner fit for NAD+ standard curve	>0.99	passed
R ² of liner fit for NADH standard curve	>0.99	passed
Volumes of single components were enough to perform two 96-well plate assays	Yes	passed

5. References

[Certificate of Analysis 0003 and 0004.docx](#)