



Title: **NADMED-0008 and -0009**

Document type: **Certificate of Analysis**

Document ID: **COA-4**

Author: Saara Forsström

Owner: Saara Forsström

Approver(s): Riikka Äänismaa
approved at 2024-04-12 14:43 (UTC +0200)

Approval date: 2024-04-12

Effective date: 2024-04-12

Next periodic review date: 2026-04-12

1. CERTIFICATE OF ANALYSIS

Product: Q-NADMED blood NAD+ and NADH assay kit, Q-NADMED blood NAD+ assay kit

Catalog numbers: IVD_001; IVD_001/TH; IVD_001_01_40; IVD_001_01_40/TH

Lots: 0008 and 0009

Specification: Performance evaluation upon manufacturing

Tested parameters:

1. UV-Vis spectra of NAD+ and/or NADH standard stocks
2. Performance of the Standards in the assay
3. Volume of individual components in the kit


Status: All parameters are within reference range


Date: 26.3.2024

Approved by Chief Scientific Officer, Liliya Euro, PhD



2. Summary of UV-Vis spectroscopy analysis of kit standards

Summary of UV-Vis spectroscopy analysis of kit standards	
Purpose: measurement of compound concentration in NAD+ and NADH standard stocks	
	<p>Standard: 1 mM NAD+</p> <p>Extinction coefficient at 260 nm: 18 mM*cm⁻¹</p> <p>Dilution: x100 with water to 10 µM</p> <p>Measurement: 1 cm quartz cuvette</p> <p>Instrument: Shimadzu UV-2401pc</p> <p>Absorbance at 260 nm: 0.183 Optical Units</p> <p>Theoretical value for absorbance of 10 µM NAD+: 0.180</p> <p>Accepted range for absorbance measurement based on instrument characteristics: 0.178 - 0.185</p> <p>Standard: 1 mM NADH</p>



Extinction coefficient at 340 nm: $6.22 \text{ mM} \cdot \text{cm}^{-1}$
 Dilution: x100 with water to 10 μM
 Measurement: 1 cm quartz cuvette
 Instrument: Shimadzu UV-2401pc
 Absorbance at 340 nm: 0.062 Optical Units
 Theoretical value for absorbance of 10 μM NADH: 0.0622
 Accepted range for absorbance measurement based on instrument characteristics: 0.059-0.066

3. Summary of Standard performance in the assays

Assay	Absorbance at 573 nm Reaction time - 4 min	Standard curve fitting																		
NAD+ Standards	<table border="1"> <thead> <tr> <th>NAD+, μM</th> <th>1 meas</th> <th>2 meas</th> </tr> </thead> <tbody> <tr> <td>0</td> <td>0.061</td> <td>0.056</td> </tr> <tr> <td>1</td> <td>0.23</td> <td>0.233</td> </tr> <tr> <td>2</td> <td>0.381</td> <td>0.391</td> </tr> <tr> <td>3</td> <td>0.536</td> <td>0.551</td> </tr> <tr> <td>5</td> <td>0.814</td> <td>0.835</td> </tr> </tbody> </table>	NAD+, μM	1 meas	2 meas	0	0.061	0.056	1	0.23	0.233	2	0.381	0.391	3	0.536	0.551	5	0.814	0.835	
	NAD+, μM	1 meas	2 meas																	
	0	0.061	0.056																	
	1	0.23	0.233																	
	2	0.381	0.391																	
	3	0.536	0.551																	
5	0.814	0.835																		
NADH Standards	<table border="1"> <thead> <tr> <th>NADH, μM</th> <th>1 meas</th> <th>2 meas</th> </tr> </thead> <tbody> <tr> <td>0</td> <td>0.058</td> <td>0.055</td> </tr> <tr> <td>0.2</td> <td>0.097</td> <td>0.094</td> </tr> <tr> <td>0.4</td> <td>0.14</td> <td>0.133</td> </tr> <tr> <td>0.6</td> <td>0.179</td> <td>0.178</td> </tr> <tr> <td>1</td> <td>0.243</td> <td>0.234</td> </tr> </tbody> </table>	NADH, μM	1 meas	2 meas	0	0.058	0.055	0.2	0.097	0.094	0.4	0.14	0.133	0.6	0.179	0.178	1	0.243	0.234	
	NADH, μM	1 meas	2 meas																	
	0	0.058	0.055																	
	0.2	0.097	0.094																	
	0.4	0.14	0.133																	
	0.6	0.179	0.178																	
1	0.243	0.234																		

4. Criteria for acceptance assay performance results

Parameter	Reference values	Quality Control
Absorbance range for 0 - 5 μ M NAD ⁺ Assay Standards (assay time 4 min)	0.04 - 0.800 Accepted variation between replicates - 0.05 Optical Units	passed
Absorbance range for 0 - 1 μ M NADH Assay Standards (assay time 6 min)	0.04 - 0.260 Accepted variation between replicates - 0.05 Optical Units	passed
R ² of liner fit for NAD ⁺ standard curve	>0.99	passed
R ² of liner fit for NADH standard curve	>0.99	passed
Volumes of single components were enough to perform two 96-well plate assays	Yes	passed

5. Attachments